6

	Application No.	Applicant(s)
		YONCE ET AL.
Notice of Allowability	10/723,255 Examiner	Art Unit
-	Lanuaged Faulcas Ir	3762
	Lenwood Faulcon, Jr.	3/62
The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.		
1. X This communication is responsive to 26 November 2003.		
2. The allowed claim(s) is/are <u>1-21</u> .		
3. The drawings filed on <u>26 November 2003</u> are accepted by the Examiner.		
4.		
Attachment(s) 1. ☑ Notice of References Cited (PTO-892) 2. ☐ Notice of Draftperson's Patent Drawing Review (PTO-948) 3. ☐ Information Disclosure Statements (PTO-1449 or PTO/SB/C Paper No./Mail Date 4. ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material	6. Interview Summary Paper No./Mail Da 7. Examiner's Amenda	te ment/Comment ent of Reasons for Allowance
	·	La Lal

Application/Control Number: 10/723,255 Page 2

Art Unit: 3762

DETAILED ACTION

Allowable Subject Matter

1. Claims 1-21 are allowed.

2. The following is an examiner's statement of reasons for allowance: claims 1-21 present allowable subject matter over the prior art since the prior art teachings do not teach of a cardiac rhythm management device and method that delivers paces to both the right and left ventricles at a programmed biventricular offset interval and after a programmed AV delay interval with respect to an atrial event, then switches to left ventricle only pacing mode while recording an evoked response test electrogram, then compares the test electrogram to an evoked response template electrogram previously recorded during left ventricle only pacing, then adjusts the programmed AV delay interval to match the patient's current intrinsic AV interval and then returns to the biventricular pacing mode with the adjusted AV delay interval.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Conclusion

3. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. Maarse (U.S. Patent No. 6,128,535), Lu (U.S. Patent No. 6,697,673), Sloman et al. (U.S. Patent No. 6,738,669), Kroll (U.S. 2001/0049543).

Application/Control Number: 10/723,255

Art Unit: 3762

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Lenwood Faulcon, Jr. whose telephone number is 571-272-6090. The examiner can normally be reached on Monday-Thursday from 9 to 5 pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Angela D. Sykes, can be reached on 571-272-4955. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Lenwood Faulcon, Jr.

George Manuel

Primary Examiner